

VITAL

Hollow Fiber Oxygenator for Adult New Ideas for Advance Perfusion

- Special Features Less priming volume Less blood damage Innovative design of heat exchanger Sterilization : ETO
- Venous Reservoir with Heat Exchanger

| Maximum volume | 4000 mL |
|--|---|
| Minimum level | 200 mL |
| Venous inlet | 12.7 mm (1 / 2 ") |
| Cardiotomy reservoir inlet | 9.53 mm (3.8") |
| Suction inlet | 2 pcs for 6.35 mm (1 / 4"), 2 pcs for 6.35 mm (1 / 4") / 9.53 mm (3 / 8") |
| Recirculation inlet | 6.35 mm (1 / 4") |
| Quick priming inlet | 6.35 mm (1 / 4") |
| Venous reservoir outlet | 9.53 mm (3 / 8") |
| Venous sample port | Luer Lock |
| Temperature probe port | Thermistor Yellow Spring type 6.35 mm (1 / 4") |
| Drug inlets | Luer lock type: 1 without filter and 3 with filter |
| Cardiotomy filter | 40 μm, 30 μm, 20 μm |
| Maximum flow in cardiotomy filter | 6L / min, 6L / min, 3L / min |
| Maximum water pressure in Heat Exchanger | 138 Kpa (20 psi) |
| Heat Exchange effective area | 0.135 m ² |
| Water inlet / outlet | Hansen type |
| Heat exchanger performance factor* | 0.6 |
| Negative / positive pressure relief valve for sealed reservoir | (-200 / +65mmHg) \pm 5 |
| | |

*4 L / min of blood flow and 10 L / min of water flow

Oxygenator Chamber

| Priming | ≅180 ml |
|------------------------|-------------------------------|
| Blood flow | 0.5 to 7.0 L / min |
| Effective area | 2.0 m ² |
| Venous inlet | 9.53 mm (3 / 8") |
| Arterial outlet | 9.53 mm (3 / 8") |
| Recirculation outlet | 6.35 mm(1 / 4") |
| Gas inlet / outlet | 6.35 mm(1 / 4") |
| Arterial sample port | Luer lock type |
| Temperature probe port | Thermistor Yellow Spring type |

Materials

| Reservoir body | Polycarbonate |
|----------------------------|--------------------|
| Filtering element | Polyester |
| Frame of filtering element | Polypropylene |
| Bubble burst filter | Polyester |
| Anti foam sponge | Polyurethane |
| Heat exchanger | Anodized aluminium |
| Oxygenator Chamber | Polycarbonate |
| Hollow fiber membrane | Polypropylene |